














Symbol Glossary

SYMBOL	SYMBOL TITLE	STANDARD REFERENCE	STANDARD TITLE	EXPLANATORY TEXT
	Catalog number	ISO 15223-1 Reference no. 5.1.6. (ISO 7000-2493)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates the manufacturer's catalog number so the medical device can be identified.
	Lot or Batch code	ISO 15223-1 Reference no. 5.1.5. (ISO 7000-2492)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates the manufacturer's batch code so the batch or lot number can be identified.
	Use by date	ISO 15223-1 Reference no. 5.1.4. (ISO 7000-2607)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates the date after which the medical device is not to be used.
	Date of manufacture	ISO 15223-1 Reference no. 5.1.3. (ISO 7000-2497)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates the date when the medical device was manufactured.
	Manufacturer	ISO 15223-1 Reference no. 5.1.1. (ISO 7000-3082)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates the medical device manufacturer.
	Sterilized using irradiation	ISO 15223-1 Reference no. 5.2.4. (ISO 7000-2502)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates a medical device that has been sterilized using irradiation.
	Do not re-use	ISO 15223-1 Reference no. 5.4.2. (ISO 7000- 1051)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates a medical device that is intended for one single use only. NOTE: Synonyms for “Do not reuse” are “single use” and “use only once”.

SYMBOL	SYMBOL TITLE	STANDARD REFERENCE	STANDARD TITLE	EXPLANATORY TEXT
	Do not use if package is damaged and consult instructions for use	ISO 15223-1 Reference no. 5.2.8. (ISO 7000-2606)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates a medical device that should not be used if the package has been damaged or opened and that the user should consult the instructions for use for additional information.
	Keep dry Keep away from rain	ISO 15223-1 Reference no. 5.3.4. (ISO 7000-0626)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates a medical device that needs protection from moisture.
	Temperature limit	ISO 15223-1 Reference no. 5.3.7. (ISO 7000-0632)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates the temperature limits to which the medical device can be safely exposed. PLEASE NOTE FOR USA ONLY: The limits 20°C to 25°C (68°F to 77°F) also mean that there are excursions permitted between 15°C to 30°C (59°F to 86°F) [See USP Controlled Room Temperature].
	Consult instructions for use	ISO 15223-1 Reference no. 5.4.3. (ISO 7000-1641)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates the need for the user to consult the instructions for use.
	Caution	ISO 15223-1 Reference no. 5.4.4. (ISO 7000-0434A)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	To indicate that caution is necessary when operating the device or control close to where the symbol is placed, or to indicate that the current situation needs operator awareness or operator action in order to avoid undesirable consequences
	Prescription Use Only	N/A	N/A	Caution: Federal law (USA) restricts this device to sale by or on the order of a licensed healthcare practitioner.

SYMBOL	SYMBOL TITLE	STANDARD REFERENCE	STANDARD TITLE	EXPLANATORY TEXT
	Quantity	N/A	N/A	Indicates the package quantity or number of medical devices included in the package.
	Heron Therapeutics Logo	N/A	N/A	Manufacturers Logo
	Do not resterilize	ISO 15223-1 Reference no. 5.2.6 (ISO 7000-2608)	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates a medical device that is not to be resterilized
	Non-Pyrogenic fluid path	ISO 15223-1:2021 Reference no. 5.6.2, 5.6.3 ISO 7000 Reference no. 2723	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	On medical devices: to indicate that the fluid path is not pyrogenic
	Consult electronic Instructions for Use	ISO 15223-1:2021 Reference no. A.16	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Example of use of symbol 5.4.3, “Consult instructions for use or consult electronic instructions for use” for an electronic instruction for use (eIFU)
	Not Made with Natural Rubber Latex	ISO 15223-1:2021 Reference 5.4.5 B2 ISO 7000 Reference no. 2725	Medical devices — Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements.	Indicates product not made with natural rubber latex
	Not Made with DEHP (Di(2- ethylhexyl) phthalate)	N/A	N/A	Indicates product does not contain DEHP